

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	01448	Calib. Date: 03/01/2016
Certification File No.:	01449	Cert. No.: 00027
Linearity File No.:	01450	Cert. Date: 03/01/2016
Solution File No.:	01516	Lin. Date: 03/01/2016
Sequential File No.:	01516	Lin. No.: 00019
		Soln. Date: 06/22/2016
		Soln. No.: 00258
		File Date: 06/22/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-113
Solution Control Lot:	14L134	Expires: 12/04/2016
		Bottle No.: 1221

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:00D		
Control 1 EC	0.100%	21:00D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	21:00D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:01D		
Control 2 EC	0.098%	21:02D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	21:02D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:02D		
Control 3 EC	0.099%	21:03D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	21:03D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:04D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: MCKAY JR

First Name: ORIN

MI: -

Signature: _____



Badge No.: 178

Date: 06/22/2016