

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01448 Calib. Date: 03/01/2016 Calib. No.: 00027
Certification File No.: 01449 Cert. Date: 03/01/2016 Cert. No.: 00019
Linearity File No.: 01450 Lin. Date: 03/01/2016 Lin. No.: 00019
Solution File No.: 01521 Soln. Date: 06/27/2016 Soln. No.: 00259
Sequential File No.: 01521 File Date: 06/27/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-113
Control Solution %: 0.100% Expires: 12/04/2016
Solution Control Lot: 14L134 Bottle No.: 1217

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:56D		
Control 1 EC	0.099%	09:57D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:57D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:58D		
Control 2 EC	0.098%	09:58D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:58D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:59D		
Control 3 EC	0.099%	09:59D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:59D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:00D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BALDINO

First Name: ANDREW

MI: M

Signature: _____



Badge No.: 174

Date: 06/27/2016