

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01448	Calib. Date:	03/01/2016
Certification File No.:	01449	Cert. Date:	03/01/2016
Linearity File No.:	01450	Lin. Date:	03/01/2016
Solution File No.:	01526	Soln. Date:	07/04/2016
Sequential File No.:	01526	File Date:	07/04/2016

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXB S3-113
Control Solution %:	0.100%			Expires:	12/04/2016
Solution Control Lot:	14L134			Bottle No.:	1222

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	00:57D		
Control 1 EC	0.099%	00:57D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	00:57D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:58D		
Control 2 EC	0.099%	00:59D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	00:59D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:59D		
Control 3 EC	0.099%	01:00D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	01:00D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:01D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DRACKWICZ

First Name: AMY

MI: M

Signature: _____

Badge No.: 177

Date: 07/04/2016