

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01109	Calib. Date:	03/01/2016
Certification File No.:	01110	Cert. Date:	03/01/2016
Linearity File No.:	01111	Lin. Date:	03/01/2016
Solution File No.:	01178	Soln. Date:	07/04/2016
Sequential File No.:	01178	File Date:	07/04/2016
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-114
Solution Control Lot:	14L134	Expires:	12/04/2016
		Bottle No.:	1211

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:58D		
Control 1 EC	0.100%	00:59D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	00:59D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:59D		
Control 2 EC	0.100%	01:00D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	01:00D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:01D		
Control 3 EC	0.100%	01:01D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	01:01D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:02D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: DRACKWICZ

First Name: AMY

MI: M

Signature: PH  177

Badge No.: 177

Date: 07/04/2016