

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	01448	Calib. Date: 03/01/2016
Certification File No.:	01449	Calib. No.: 00027
Linearity File No.:	01450	Cert. Date: 03/01/2016
Solution File No.:	01532	Cert. No.: 00019
Sequential File No.:	01532	Lin. Date: 03/01/2016
		Lin. No.: 00019
		Soln. Date: 07/17/2016
		Soln. No.: 00262
		File Date: 07/17/2016

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-113
Solution Control Lot:	201603L	Expires: 03/22/2018
		Bottle No.: 0845

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:08D		
Control 1. EC	0.101%	07:08D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	07:08D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:09D		
Control 2 EC	0.100%	07:09D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	07:09D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:10D		
Control 3 EC	0.100%	07:11D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	07:11D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:11D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DRACKWICZ

First Name: AMY

MI: M

Signature:  177

Badge No.: 177

Date: 07/17/2016