

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01109 Calib. Date: 03/01/2016 Calib. No.: 00020
Certification File No.: 01110 Cert. Date: 03/01/2016 Cert. No.: 00017
Linearity File No.: 01111 Lin. Date: 03/01/2016 Lin. No.: 00017
Solution File No.: 01188 Soln. Date: 07/17/2016 Soln. No.: 00222
Sequential File No.: 01188 File Date: 07/17/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-114
Control Solution %: 0.100% Expires: 03/22/2018
Solution Control Lot: 201603L Bottle No.: 0842

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	04:15D		
Control 1 EC	0.103%	04:16D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	04:16D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:16D		
Control 2 EC	0.102%	04:17D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.104%	04:17D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:17D		
Control 3 EC	0.102%	04:18D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	04:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:19D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DRACKWICZ

First Name: AMY

MI: M

Signature: _____



Badge No.: 177

Date: 07/17/2016