

Calibrating Unit New Standard Solution Report

| | | | |
|-------------------------|-----------------------|--------------|-------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: | ARXA-0037 |
| Location: | WALL TOWNSHIP POLICE | | |
| Calibration File No.: | 01448 | Calib. Date: | 03/01/2016 |
| Certification File No.: | 01449 | Cert. Date: | 03/01/2016 |
| Linearity File No.: | 01450 | Lin. Date: | 03/01/2016 |
| Solution File No.: | 01539 | Soln. Date: | 08/01/2016 |
| Sequential File No.: | 01539 | File Date: | 08/01/2016 |
| | | | |
| Calibrating Unit: | WET | Model No.: | CU-34 |
| Control Solution %: | 0.100% | Serial No.: | DDXB S3-113 |
| Solution Control Lot: | 14L134 | Expires: | 12/04/2016 |
| | | Bottle No.: | 1216 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 02:38D | | |
| Control 1 EC | 0.099% | 02:39D | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.099% | 02:39D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 02:39D | | |
| Control 2 EC | 0.099% | 02:40D | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 02:40D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 02:41D | | |
| Control 3 EC | 0.098% | 02:41D | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 02:41D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 02:42D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DRACKWICZ

First Name: AMY

MI: M

Signature: _____

PH  *177*

Badge No.: 177

Date: 08/01/2016