

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	01109	Calib. Date: 03/01/2016
Certification File No.:	01110	Calib. No.: 00020
Linearity File No.:	01111	Cert. Date: 03/01/2016
Solution File No.:	01196	Cert. No.: 00017
Sequential File No.:	01196	Lin. Date: 03/01/2016
		Lin. No.: 00017
		Soln. Date: 08/03/2016
		Soln. No.: 00223
		File Date: 08/03/2016

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-114
Solution Control Lot:	201603L	Expires: 03/22/2018
		Bottle No.: 0840

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:50D		
Control 1 EC	0.100%	09:51D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	09:51D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:51D		
Control 2 EC	0.100%	09:52D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.103%	09:52D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:53D		
Control 3 EC	0.101%	09:53D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.102%	09:53D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:54D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BALDINO

First Name: ANDREW

MI: M

Signature: _____



Badge No.: 174

Date: 08/03/2016