

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01213 Calib. Date: 09/15/2016 Calib. No.: 00025
Certification File No.: 01214 Cert. Date: 09/15/2016 Cert. No.: 00021
Linearity File No.: 01215 Lin. Date: 09/15/2016 Lin. No.: 00021
Solution File No.: 01218 Soln. Date: 09/28/2016 Soln. No.: 00226
Sequential File No.: 01218 File Date: 09/28/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAB-0015
Control Solution %: 0.100% Expires: 03/22/2018
Solution Control Lot: 201603L Bottle No.: 0843

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:29D		
Control 1 EC	0.099%	07:30D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	07:30D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:30D		
Control 2 EC	0.099%	07:31D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	07:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:31D		
Control 3 EC	0.100%	07:32D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	07:32D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:33D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCKAY JR

First Name: ORIN

MI: -

Signature: _____



Badge No.: 178

Date: 09/28/2016