

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	01553	Calib. Date: 09/15/2016
Certification File No.:	01554	Calib. No.: 00031
Linearity File No.:	01555	Cert. Date: 09/15/2016
Solution File No.:	01571	Cert. No.: 00022
Sequential File No.:	01571	Lin. Date: 09/15/2016
		Lin. No.: 00022
		Soln. Date: 10/16/2016
		Soln. No.: 00269
		File Date: 10/16/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAB-0014
Solution Control Lot:	201605E	Expires: 05/17/2018
		Bottle No.: 0029

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:44D		
Control 1 EC	0.099%	02:45D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	02:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:45D		
Control 2 EC	0.098%	02:46D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	02:46D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:47D		
Control 3 EC	0.099%	02:47D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	02:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:48D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCKAY JR

First Name: ORIN

MI: -

Signature: _____



Badge No.: 178

Date: 10/16/2016