

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01553	Calib. Date:	09/15/2016
Certification File No.:	01554	Cert. Date:	09/15/2016
Linearity File No.:	01555	Lin. Date:	09/15/2016
Solution File No.:	01577	Soln. Date:	10/27/2016
Sequential File No.:	01577	File Date:	10/27/2016
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0014
Solution Control Lot:	201605E	Expires:	05/17/2018
		Bottle No.:	0030

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:02D		
Control 1 EC	0.100%	10:02D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	10:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:03D		
Control 2 EC	0.098%	10:03D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:04D		
Control 3 EC	0.098%	10:05D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:05D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:05D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL.

MI: A

Signature: _____

PTL. K #129

Badge No.: 129

Date: 10/27/2016