

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01213	Calib. Date:	09/15/2016
Certification File No.:	01214	Cert. Date:	09/15/2016
Linearity File No.:	01215	Lin. Date:	09/15/2016
Solution File No.:	01226	Soln. Date:	10/27/2016
Sequential File No.:	01226	File Date:	10/27/2016
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0015
Solution Control Lot:	201605E	Expires:	05/17/2018
		Bottle No.:	0045

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:07D		
Control 1 EC	0.100%	10:08D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:08D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:08D		
Control 2 EC	0.100%	10:09D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:09D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:09D		
Control 3 EC	0.100%	10:10D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	10:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:11D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL.

MI: A

Signature: _____



Badge No.: 129

Date: 10/27/2016