

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01553 Calib. Date: 09/15/2016 Calib. No.: 00031
Certification File No.: 01554 Cert. Date: 09/15/2016 Cert. No.: 00022
Linearity File No.: 01555 Lin. Date: 09/15/2016 Lin. No.: 00022
Solution File No.: 01588 Soln. Date: 11/24/2016 Soln. No.: 00271
Sequential File No.: 01588 File Date: 11/24/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAB-0014
Control Solution %: 0.100% Expires: 05/17/2018
Solution Control Lot: 201605E Bottle No.: 0050

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:17S		
Control 1 EC	0.099%	06:18S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	06:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:18S		
Control 2 EC	0.099%	06:19S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	06:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:20S		
Control 3 EC	0.098%	06:20S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	06:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:21S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DRACKWICZ

First Name: AMY

MI: M

Signature: _____



Badge No.: 177

Date: 11/24/2016