

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	01213	Calib. Date: 09/15/2016
Certification File No.:	01214	Calib. No.: 00025
Linearity File No.:	01215	Cert. Date: 09/15/2016
Solution File No.:	01234	Cert. No.: 00021
Sequential File No.:	01234	Lin. Date: 09/15/2016
		Lin. No.: 00021
		Soln. Date: 11/24/2016
		Soln. No.: 00228
		File Date: 11/24/2016

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDAB-0015
Control Solution %:	0.100%		Expires: 05/17/2018
Solution Control Lot:	201605E		Bottle No.: 0046

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	06:21S		
Control 1 EC	0.096%	06:22S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.096%	06:22S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:22S		
Control 2 EC	0.096%	06:23S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	06:23S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:24S		
Control 3 EC	0.096%	06:24S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	06:24S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:25S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DRACKWICZ

First Name: AMY

MI: M

Signature:  177

Badge No.: 177

Date: 11/24/2016