

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01553	Calib. Date:	09/15/2016
Certification File No.:	01554	Cert. Date:	09/15/2016
Linearity File No.:	01555	Lin. Date:	09/15/2016
Solution File No.:	01598	Soln. Date:	12/15/2016
Sequential File No.:	01598	File Date:	12/15/2016
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0014
Solution Control Lot:	201605E	Expires:	05/17/2018
		Bottle No.:	0047

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:07S		
Control 1 EC	0.099%	08:08S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:08S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:09S		
Control 2 EC	0.097%	08:09S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	08:09S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:10S		
Control 3 EC	0.097%	08:10S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	08:10S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:11S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: DRACKWICZ

First Name: AMY

MI: M

Signature: 

Badge No.: 177

Date: 12/15/2016