

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01213 Calib. Date: 09/15/2016 Calib. No.: 00025
Certification File No.: 01214 Cert. Date: 09/15/2016 Cert. No.: 00021
Linearity File No.: 01215 Lin. Date: 09/15/2016 Lin. No.: 00021
Solution File No.: 01239 Soln. Date: 12/15/2016 Soln. No.: 00229
Sequential File No.: 01239 File Date: 12/15/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAB-0015
Control Solution %: 0.100% Expires: 05/17/2018
Solution Control Lot: 201605E Bottle No.: 0044

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:14S		
Control 1 EC	0.099%	08:15S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:15S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:16S		
Control 2 EC	0.099%	08:16S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:16S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:17S		
Control 3 EC	0.100%	08:18S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	08:18S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:18S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DRACKWICZ

First Name: AMY

MI: M

Signature: 

Badge No.: 177

Date: 12/15/2016