

Calibrating Unit New Standard Solution Report


Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01553	Calib. Date:	09/15/2016
Certification File No.:	01554	Cert. Date:	09/15/2016
Linearity File No.:	01555	Lin. Date:	09/15/2016
Solution File No.:	01605	Soln. Date:	12/30/2016
Sequential File No.:	01605	File Date:	12/30/2016
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0014
Solution Control Lot:	201605E	Expires:	05/17/2018
		Bottle No.:	0032

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:12S		
Control 1 EC	0.098%	08:12S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:13S		
Control 2 EC	0.097%	08:13S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:14S		
Control 3 EC	0.098%	08:15S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:15S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY	First Name: PTL.	MI: A
Signature: 		Badge No.: 129
		Date: 12/30/2016