

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01553	Calib. Date:	09/15/2016
Certification File No.:	01554	Cert. No.:	00031
Linearity File No.:	01555	Cert. Date:	09/15/2016
Solution File No.:	01609	Lin. Date:	09/15/2016
Sequential File No.:	01609	Lin. No.:	00022
		Soln. Date:	01/13/2017
		Soln. No.:	00275
		File Date:	01/13/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0014
Solution Control Lot:	201605E	Expires:	05/17/2018
		Bottle No.:	0028

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:57S		
Control 1 EC	0.100%	08:58S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:59S		
Control 2 EC	0.098%	08:59S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:00S		
Control 3 EC	0.099%	09:01S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:01S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL.

MI: A

Signature: _____

PTL K #129

Badge No.: 129

Date: 01/13/2017