

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARZL-0146  
Location: WALL TOWNSHIP POLICE  
Calibration File No.: 01213 Calib. Date: 09/15/2016 Calib. No.: 00025  
Certification File No.: 01214 Cert. Date: 09/15/2016 Cert. No.: 00021  
Linearity File No.: 01215 Lin. Date: 09/15/2016 Lin. No.: 00021  
Solution File No.: 01253 Soln. Date: 01/24/2017 Soln. No.: 00232  
Sequential File No.: 01253 File Date: 01/24/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAB-0015  
Control Solution %: 0.100% Expires: 05/17/2018  
Solution Control Lot: 201605E Bottle No.: 0042

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:33S		
Control 1 EC	0.096%	21:34S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	21:34S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:35S		
Control 2 EC	0.096%	21:35S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	21:35S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:36S		
Control 3 EC	0.097%	21:37S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	21:37S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:38S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


## Changed By:

Last Name: SANTORO

First Name: DANIEL

MI: J

Signature: \_\_\_\_\_



Badge No.: 165

Date: 01/24/2017