

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01613	Calib. Date:	01/25/2017
Certification File No.:	01614	Cert. Date:	01/25/2017
Linearity File No.:	01615	Lin. Date:	01/25/2017
Solution File No.:	01617	Soln. Date:	02/10/2017
Sequential File No.:	01617	File Date:	02/10/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-113
Solution Control Lot:	201605E	Expires:	05/17/2018
		Bottle No.:	0043

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:45S		
Control 1 EC	0.101%	08:46S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:46S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:46S		
Control 2 EC	0.100%	08:47S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:47S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:48S		
Control 3 EC	0.100%	08:48S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	08:48S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:49S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL.

MI: A.

Signature: _____

PTL. K #129

Badge No.: 129

Date: 02/10/2017