

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	01254	Calib. Date: 01/25/2017
Certification File No.:	01255	Calib. No.: 00026
Linearity File No.:	01256	Cert. Date: 01/25/2017
Solution File No.:	01260	Cert. No.: 00022
Sequential File No.:	01260	Lin. Date: 01/25/2017
		Lin. No.: 00022
		Soln. Date: 02/10/2017
		Soln. No.: 00234
		File Date: 02/10/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-114
Solution Control Lot:	16330	Expires: 10/28/2018
		Bottle No.: 0295

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:51S		
Control 1 EC	0.100%	08:52S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:52S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:52S		
Control 2 EC	0.099%	08:53S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:53S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:54S		
Control 3 EC	0.100%	08:54S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:54S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:55S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL.

MI: A.

Signature: _____

PTL. K #129

Badge No.: 129

Date: 02/10/2017