

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Wall Twp Police Department 2700 Allaire Road Wall, NJ 07719	2. Contact: Todd Kazanowsky 3. Phone Number: 732-449-4500	4. Date: 10/30/19 5. County: Monmouth
6. Alcotest Instrument Serial Number: ARZL-0146	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A

9. Reason for Service:
Complaint received of the instrument experiencing multiple control test failures. Complaint received of error message "075 HEATER SYSTEM".

10. Comments:
Operational condition of the instrument was assessed by reviewing the performance history of the instrument. This revealed multiple control test failures due to low electrochemical sensor (EC) results. The instrument was placed out of service and returned to Draeger Safety Diagnostics for evaluation.

Last Known Sequential File #: 1765

☐ 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.

☒ 12. The above Instrument/Component is placed out of service pending further evaluation.

☐ 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I Dennis J. Lutz #7045
Name & Badge Number (Print)

Tpr I Lutz 7045 10-30-19
Signature Date

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Wall Twp Police Department 2700 Allaire Road Wall, NJ 07719		2. Contact: Todd Kazanowsky	4. Date: 12/13/19
		3. Phone Number: 732-449-4500	5. County: Monmouth
6. Alcotest Instrument Serial Number: ARZL-0146	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A	

9. Reason for Service:

The alcotest instrument was returned from outside evaluation and placed back in service.

10. Comments:

See Draeger Return and Repair form.

Last Known Sequential File #: NA

☐ 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.

☐ 12. The above Instrument/Component is placed out of service pending further evaluation.

☒ 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I Dennis J. Lutz #7045
Name & Badge Number (Print)

Signature

Tpr I Lutz 7045 12-13-19
Date

RETURN AND REPAIR FORM

Dräger

Customer Information

B:

S:

Company Name: Wall Township Police, Wall, New Jersey

Date Received: 11/05/19

Date given to service: 11/05/19

Carrier: ☐ FedEx ☐ UPS ☐ USPS Shipping Method: ☐ GRD ☐ 3DAY ☐ 2DAY
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810
☒ 7110 ☐ 8610
☐ 7510 ☐ DT5000
☐ 7410 Upper-half ☐ 9510
☐ 7410 Whole

Serial Number: AR ZL-0146

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case
☐ Dry Gas ☐ Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	
MP Labor	Labor	1	
6808455	Fuel Cell	1	

Repair Notes: Replaced defective fuel cell.

Customer complaint of receiving Error 75. Ran tests to duplicate Error 75 and was unable to duplicate.
Blew out cuvette of debris.

FUEL CELL = ARMH-0511

CAL W/QC AND OPS CHECK

Service Technician: BS

Date: 12/03/19

Delivery receipt



Customer no.
150062190

Number of report
304694235
Please reference on inquiries

Date of receiving	Delivery receipt date
/ /	12/03/2019

Consignee, 150062190
TOWNSHIP OF WALL
ATTN: POLICE DEPARTMENT
2700 ALLAIRE RD
WALL TOWNSHIP NJ 07719-9570

Customer, 150062190
TOWNSHIP OF WALL
ATTN: POLICE DEPARTMENT
2700 ALLAIRE RD
WALL TOWNSHIP NJ 07719-9570

Your order

Date of order:
11/12/2019
Purch.ord.no.:
19-05382

Branch text

Draeger contact person

Service Support-IDT
800-437-2437 OPT 5
SS-IDT@draeger.com

Page 1

Quantity	Description	Part no.	Equipment - No.	Serial - No.
001	ALCOTEST 7110 MK III USA	8314246	1029079772	ARZL-0146

Draeger Inc.
Our Tax ID: 23-1699096
3135 Quarry Road; Telford, PA 18969
An Equal Opportunity Employer M / F / V / H
Telephone 800-437-2437
<http://www.draeger.com>

Remit to:
LOCKBOX (Standard USPS)
Draeger, Inc.
PO Box 13369
Newark, New Jersey
07101-3362

Remit to:
LOCKBOX (Overnight)
FIS Lockbox Processing
Lockbox #13369
100 Grove Road
Suite E
West Deptford, NJ 08066

Remit US Wire Transfers to:
Account Name: Draeger Inc.
Account Number: 00-494-936
Transit Routing: 021001033
SWIFT: BKTRUS33
Deutsche Bank Trust Company Americas
60 Wall Street 25th Fl, New York, NY 10005

Service note

Customer no. 150062190
Number of report 304694235
Date of receiving 10/31/2019
Please reference on inquiries

Customer
TOWNSHIP OF WALL
ATTN: POLICE DEPARTMENT
2700 ALLAIRE RD
WALL TOWNSHIP NJ 07719-9570

Consignee
TOWNSHIP OF WALL
ATTN: POLICE DEPARTMENT
2700 ALLAIRE RD
WALL TOWNSHIP NJ 07719-9570

Your order

Date of order:
11/12/2019
Your reference:
19-05382

NL-Text

Maintenance device list

Your contact person

Service Support-IDT
800-437-2437 OPT 5
SS-IDT@draeger.com

Page 1

Item	Quantity	Part no. Description
		Service Order: 142021399
		COE005615
		Eq.No. : 1029079772 SN: ARZL-0146
		Material: 8314246 ALCOTEST 7110 MK III USA
		Inv.No. :
		Location:
		COE005615
	1.0 H	R018 Repair time - Core
	1 EA	6808455 Sensor Alcotest B

Service note

Customer no. 150062190
Number of report 304694235
Date of receiving 10/31/2019
Please reference on inquiries

Customer
TOWNSHIP OF WALL
ATTN: POLICE DEPARTMENT
2700 ALLAIRE RD
WALL TOWNSHIP NJ 07719-9570

Consignee
TOWNSHIP OF WALL
ATTN: POLICE DEPARTMENT
2700 ALLAIRE RD
WALL TOWNSHIP NJ 07719-9570

Your order

Date of order:
11/12/2019
Your reference:
19-05382

NL-Text

Maintenance device list

Your contact person

Service Support-IDT
800-437-2437 OPT 5
SS-IDT@draeger.com

Page 1

Item	Quantity	Part no. Description
		Service Order: 142021399
		COE005615
		Eq.No. : 1029079772 SN: ARZL-0146
		Material: 8314246 ALCOTEST 7110 MK III USA
		Inv.No. :
		Location:
		COE005615
1.0	H	R018 Repair time - Core
1	EA	6808455 Sensor Alcotest B