

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00225	Calib. Date:	09/29/2010
Certification File No.:	00226	Cert. Date:	09/29/2010
Linearity File No.:	00227	Lin. Date:	09/29/2010
Solution File No.:	00277	Soln. Date:	02/16/2011
Sequential File No.:	00277	File Date:	02/16/2011
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0014
Solution Control Lot:	10F079	Expires:	06/10/2012
		Bottle No.:	0381

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:38S		
Control 1 EC	0.102%	10:39S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:39S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:40S		
Control 2 EC	0.096%	10:40S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.095%	10:40S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:41S		
Control 3 EC	0.095%	10:42S	34.0°C	CONTROL TEST FAILED
Control 3 IR	0.093%	10:42S	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	10:43S		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

Badge No.: 129

Date: 02/16/2011