

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00494	Calib. Date:	09/29/2011
Certification File No.:	00495	Cert. Date:	09/29/2011
Linearity File No.:	00496	Lin. Date:	09/29/2011
Solution File No.:	00504	Soln. Date:	11/04/2011
Sequential File No.:	00504	File Date:	11/04/2011
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0014
Solution Control Lot:	10L084	Expires:	12/16/2012
		Bottle No.:	0764

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:22S		
Control 1 EC	0.099%	07:22S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	07:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:23S		
Control 2 EC	0.099%	07:24S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	07:24S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:24S		
Control 3 EC	0.099%	07:25S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	07:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:25S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

Badge No.: 129

Date: 11/04/2011