

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 00494 Calib. Date: 09/29/2011 Calib. No.: 00014
Certification File No.: 00495 Cert. Date: 09/29/2011 Cert. No.: 00008
Linearity File No.: 00496 Lin. Date: 09/29/2011 Lin. No.: 00008
Solution File No.: 00508 Soln. Date: 11/18/2011 Soln. No.: 00103
Sequential File No.: 00508 File Date: 11/18/2011

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAB-0014
Control Solution %: 0.100% Expires: 03/31/2013
Solution Control Lot: 11C088 Bottle No.: 0521

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:33S		
Control 1 EC	0.096%	08:34S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	08:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:34S		
Control 2 EC	0.097%	08:35S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:35S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:35S		
Control 3 EC	0.098%	08:36S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	08:36S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:37S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: PTL. K. 129

Badge No.: 129

Date: 11/18/2011