

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 00459 **Calib. Date:** 03/07/2012 **Calib. No.:** 00010
Certification File No.: 00460 **Cert. Date:** 03/07/2012 **Cert. No.:** 00008
Linearity File No.: 00461 **Lin. Date:** 03/07/2012 **Lin. No.:** 00008
Solution File No.: 00497 **Soln. Date:** 07/06/2012 **Soln. No.:** 00091
Sequential File No.: 00497 **File Date:** 07/06/2012

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDXB S3-114
Control Solution %: 0.100% **Expires:** 08/17/2013
Solution Control Lot: 11H094 **Bottle No.:** 0317

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:21D		
Control 1 EC	0.101%	03:21D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	03:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:22D		
Control 2 EC	0.099%	03:23D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	03:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:23D		
Control 3 EC	0.099%	03:24D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	03:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:25D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: COSTANTINI

First Name: JASON

MI: R

Signature: 

Badge No.: 152

Date: 07/06/2012