

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 00562 **Calib. Date:** 03/01/2013 **Calib. No.:** 00012
Certification File No.: 00563 **Cert. Date:** 03/01/2013 **Cert. No.:** 00010
Linearity File No.: 00564 **Lin. Date:** 03/01/2013 **Lin. No.:** 00010
Solution File No.: 00584 **Soln. Date:** 04/12/2013 **Soln. No.:** 00111
Sequential File No.: 00584 **File Date:** 04/12/2013

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDXB S3-113
Control Solution %: 0.100% **Expires:** 08/17/2013
Solution Control Lot: 11H094 **Bottle No.:** 0305

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:18D		
Control 1 EC	0.099%	08:19D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	08:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:20D		
Control 2 EC	0.096%	08:20D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:21D		
Control 3 EC	0.097%	08:22D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	08:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:23D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: PTL KAZANOWSKY

Badge No.: 129

Date: 04/12/2013