

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 00562 Calib. Date: 03/01/2013 Calib. No.: 00012
Certification File No.: 00563 Cert. Date: 03/01/2013 Cert. No.: 00010
Linearity File No.: 00564 Lin. Date: 03/01/2013 Lin. No.: 00010
Solution File No.: 00589 Soln. Date: 05/17/2013 Soln. No.: 00114
Sequential File No.: 00589 File Date: 05/17/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-113
Control Solution %: 0.100% Expires: 08/17/2013
Solution Control Lot: 11H094 Bottle No.: 0303

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:14D		
Control 1 EC	-.---%	14:15D	22.8°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	14:15D	22.8°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

Badge No.: 129

Date: 05/17/2013