

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: WALL TOWNSHIP POLICE
Serial No.: ARZL-0146
Calibration File No.: 00562
Calib. Date: 03/01/2013
Calib. No.: 00012
Certification File No.: 00563
Cert. Date: 03/01/2013
Cert. No.: 00010
Linearity File No.: 00564
Lin. Date: 03/01/2013
Lin. No.: 00010
Solution File No.: 00649
Soln. Date: 08/19/2013
Soln. No.: 00126
Sequential File No.: 00649
File Date: 08/19/2013

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 12F102
Model No.: CU-34
Serial No.: DDXB S3-113
Expires: 06/28/2014
Bottle No.: 0916

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	00:30D		
Control 1 EC	0.098%	00:31D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	00:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:32D		
Control 2 EC	0.097%	00:32D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	00:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:33D		
Control 3 EC	0.097%	00:34D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	00:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:35D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KELLY

First Name: JASON

MI: R

Signature: 

Badge No.: 166

Date: 08/19/2013

*e-mailed to #157
8-20-13*