

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 00655 **Calib. Date:** 08/28/2013 **Calib. No.:** 00013
Certification File No.: 00656 **Cert. Date:** 08/28/2013 **Cert. No.:** 00011
Linearity File No.: 00657 **Lin. Date:** 08/28/2013 **Lin. No.:** 00011
Solution File No.: 00680 **Soln. Date:** 10/31/2013 **Soln. No.:** 00134
Sequential File No.: 00680 **File Date:** 10/31/2013

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDAB-0014
Control Solution %: 0.100% **Expires:** 06/17/2015
Solution Control Lot: 13F114 **Bottle No.:** 1113

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:10S		
Control 1 EC	-.---%	08:10S	23.3°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	08:10S	23.3°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

Badge No.: 129

Date: 10/31/2013