

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 00655 **Calib. Date:** 08/28/2013 **Calib. No.:** 00013
Certification File No.: 00656 **Cert. Date:** 08/28/2013 **Cert. No.:** 00011
Linearity File No.: 00657 **Lin. Date:** 08/28/2013 **Lin. No.:** 00011
Solution File No.: 00694 **Soln. Date:** 12/27/2013 **Soln. No.:** 00140
Sequential File No.: 00694 **File Date:** 12/27/2013

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDAB-0014
Control Solution %: 0.100% **Expires:** 07/18/2015
Solution Control Lot: 13G115 **Bottle No.:** 1210

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	01:54S		
Control 1 EC	0.099%	01:55S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	01:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:56S		
Control 2 EC	0.096%	01:57S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	01:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:58S		
Control 3 EC	0.097%	01:58S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	01:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:59S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CADIGAN

First Name: JAMES

MI: R

Signature: _____

Badge No.: 162

Date: 12/27/2013

*e-mailed to Lauber
1-2-14*