

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 00899 **Calib. Date:** 08/28/2013 **Calib. No.:** 00020
Certification File No.: 00900 **Cert. Date:** 08/28/2013 **Cert. No.:** 00013
Linearity File No.: 00901 **Lin. Date:** 08/28/2013 **Lin. No.:** 00013
Solution File No.: 00998 **Soln. Date:** 01/30/2014 **Soln. No.:** 00166
Sequential File No.: 00998 **File Date:** 01/30/2014

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDAB-0015
Control Solution %: 0.100% **Expires:** 06/17/2015
Solution Control Lot: 13F114 **Bottle No.:** 1125

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:17S		
Control 1 EC	0.100%	11:17S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:18S		
Control 2 EC	0.099%	11:18S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:19S		
Control 3 EC	0.099%	11:20S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:21S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PTL #129

Badge No.: 129

Date: 01/30/2014

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 00655 **Calib. Date:** 08/28/2013 **Calib. No.:** 00013
Certification File No.: 00656 **Cert. Date:** 08/28/2013 **Cert. No.:** 00011
Linearity File No.: 00657 **Lin. Date:** 08/28/2013 **Lin. No.:** 00011
Solution File No.: 00708 **Soln. Date:** 01/30/2014 **Soln. No.:** 00147
Sequential File No.: 00708 **File Date:** 01/30/2014

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDAB-0014
Control Solution %: 0.100% **Expires:** 07/18/2015
Solution Control Lot: 13G115 **Bottle No.:** 1214

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:26S		
Control 1 EC	0.100%	11:27S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:28S		
Control 2 EC	0.096%	11:29S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:30S		
Control 3 EC	0.097%	11:30S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:31S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: PTL KAZANOWSKY

Badge No.: 129

Date: 01/30/2014