

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01004 **Calib. Date:** 02/10/2014 **Calib. No.:** 00021
Certification File No.: 01005 **Cert. Date:** 02/10/2014 **Cert. No.:** 00014
Linearity File No.: 01006 **Lin. Date:** 02/10/2014 **Lin. No.:** 00014
Solution File No.: 01046 **Soln. Date:** 03/27/2014 **Soln. No.:** 00176
Sequential File No.: 01046 **File Date:** 03/27/2014

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDXB S3-113
Control Solution %: 0.100% **Expires:** 07/18/2015
Solution Control Lot: 13G115 **Bottle No.:** 1212

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:56S		
Control 1 EC	0.098%	08:57S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	08:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:57S		
Control 2 EC	0.096%	08:58S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	08:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:59S		
Control 3 EC	0.096%	08:59S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	08:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:00S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL.

MI: A

Signature: PTL. K #129

Badge No.: 129

Date: 03/27/2014

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01004 **Calib. Date:** 02/10/2014 **Calib. No.:** 00021
Certification File No.: 01005 **Cert. Date:** 02/10/2014 **Cert. No.:** 00014
Linearity File No.: 01006 **Lin. Date:** 02/10/2014 **Lin. No.:** 00014
Solution File No.: 01045 **Soln. Date:** 03/27/2014 **Soln. No.:** 00175
Sequential File No.: 01045 **File Date:** 03/27/2014

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDXB S3-113
Control Solution %: 0.100% **Expires:** 07/18/2015
Solution Control Lot: 13G115 **Bottle No.:** 1212

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:31S		
Control 1 EC	-.---%	07:31S	21.1°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	07:31S	21.1°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: PTL #129

Badge No.: 129

Date: 03/27/2014