

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01062 **Calib. Date:** 05/06/2014 **Calib. No.:** 00023
Certification File No.: 01063 **Cert. Date:** 05/06/2014 **Cert. No.:** 00015
Linearity File No.: 01064 **Lin. Date:** 05/06/2014 **Lin. No.:** 00015
Solution File No.: 01073 **Soln. Date:** 05/14/2014 **Soln. No.:** 00182
Sequential File No.: 01073 **File Date:** 05/14/2014

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDXB S3-113
Control Solution %: 0.100% **Expires:** 07/18/2015
Solution Control Lot: 13G115 **Bottle No.:** 1202

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:56D		
Control 1 EC	0.097%	09:56D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.096%	09:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:57D		
Control 2 EC	0.097%	09:57D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	09:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:58D		
Control 3 EC	0.097%	09:58D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	09:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:59D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: PTL. K #129

Badge No.: 129

Date: 05/14/2014