

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 00722 **Calib. Date:** 04/11/2014 **Calib. No.:** 00016
Certification File No.: 00723 **Cert. Date:** 04/11/2014 **Cert. No.:** 00013
Linearity File No.: 00724 **Lin. Date:** 04/11/2014 **Lin. No.:** 00013
Solution File No.: 00771 **Soln. Date:** 07/05/2014 **Soln. No.:** 00158
Sequential File No.: 00771 **File Date:** 07/05/2014

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDXB S3-114
Control Solution %: 0.100% **Expires:** 09/10/2015
Solution Control Lot: 131119 **Bottle No.:** 0278

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:26D		
Control 1 EC	0.099%	14:27D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	14:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:27D		
Control 2 EC	0.098%	14:28D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:28D		
Control 3 EC	0.098%	14:29D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:30D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with
Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BALDINO

First Name: ANDREW

MI: M

Signature:  #174

Badge No.: 174

Date: 07/05/2014