

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0037  
Location: WALL TOWNSHIP POLICE  
Calibration File No.: 01062 Calib. Date: 05/06/2014 Calib. No.: 00023  
Certification File No.: 01063 Cert. Date: 05/06/2014 Cert. No.: 00015  
Linearity File No.: 01064 Lin. Date: 05/06/2014 Lin. No.: 00015  
Solution File No.: 01155 Soln. Date: 08/08/2014 Soln. No.: 00192  
Sequential File No.: 01155 File Date: 08/08/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-113  
Control Solution %: 0.100% Expires: 09/10/2015  
Solution Control Lot: 131119 Bottle No.: 0282

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:19D		
Control 1 EC	0.098%	06:20D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	06:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:21D		
Control 2 EC	0.097%	06:21D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	06:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:22D		
Control 3 EC	0.097%	06:22D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	06:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:23D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BALDINO

First Name: ANDREW

MI: M

Signature: \_\_\_\_\_

Badge No.: 174

Date: 08/08/2014