

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01191 **Calib. Date:** 10/10/2014 **Calib. No.:** 00024
Certification File No.: 01192 **Cert. Date:** 10/10/2014 **Cert. No.:** 00016
Linearity File No.: 01193 **Lin. Date:** 10/10/2014 **Lin. No.:** 00016
Solution File No.: 01213 **Soln. Date:** 11/24/2014 **Soln. No.:** 00200
Sequential File No.: 01213 **File Date:** 11/24/2014

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDAB-0014
Control Solution %: 0.100% **Expires:** 04/29/2016
Solution Control Lot: 14D129 **Bottle No.:** 0147

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:23S		
Control 1 EC	0.100%	03:24S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	03:24S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:24S		
Control 2 EC	0.099%	03:25S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	03:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:26S		
Control 3 EC	0.100%	03:26S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	03:26S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:27S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: OHALLORAN

First Name: SEAN

MI: P

Signature: _____

Badge No.: 146

Date: 11/24/2014