

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0037  
**Location:** WALL TOWNSHIP POLICE  
**Calibration File No.:** 01191 **Calib. Date:** 10/10/2014 **Calib. No.:** 00024  
**Certification File No.:** 01192 **Cert. Date:** 10/10/2014 **Cert. No.:** 00016  
**Linearity File No.:** 01193 **Lin. Date:** 10/10/2014 **Lin. No.:** 00016  
**Solution File No.:** 01224 **Soln. Date:** 12/30/2014 **Soln. No.:** 00203  
**Sequential File No.:** 01224 **File Date:** 12/30/2014

**Calibrating Unit:** WET **Model No.:** CU-34 **Serial No.:** DDAB-0014  
**Control Solution %:** 0.100% **Expires:** 04/29/2016  
**Solution Control Lot:** 14D129 **Bottle No.:** 0136

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 11:48S        |                               |                           |
| Control 1 EC      | 0.099%         | 11:49S        | 33.9°C                        | *** TEST PASSED ***       |
| Control 1 IR      | 0.099%         | 11:49S        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 11:49S        |                               |                           |
| Control 2 EC      | 0.099%         | 11:50S        | 34.0°C                        | *** TEST PASSED ***       |
| Control 2 IR      | 0.099%         | 11:50S        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 11:51S        |                               |                           |
| Control 3 EC      | 0.099%         | 11:51S        | 33.9°C                        | *** TEST PASSED ***       |
| Control 3 IR      | 0.099%         | 11:51S        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 11:52S        |                               |                           |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: KAZANOWSKY

First Name: PTL.

MI: A

Signature: PTL KAZANOWSKY

Badge No.: 129

Date: 12/30/2014