

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01259 **Calib. Date:** 03/25/2015 **Calib. No.:** 00025
Certification File No.: 01260 **Cert. Date:** 03/25/2015 **Cert. No.:** 00017
Linearity File No.: 01261 **Lin. Date:** 03/25/2015 **Lin. No.:** 00017
Solution File No.: 01362 **Soln. Date:** 09/06/2015 **Soln. No.:** 00222
Sequential File No.: 01362 **File Date:** 09/06/2015

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDXB S3-113
Control Solution %: 0.100% **Expires:** 06/13/2017
Solution Control Lot: 15F139 **Bottle No.:** 0479

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	04:53D		
Control 1 EC	0.097%	04:54D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	04:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:54D		
Control 2 EC	0.096%	04:55D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	04:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:56D		
Control 3 EC	0.096%	04:56D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	04:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:57D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BALDINO

First Name: ANDREW

MI: M

Signature: 

Badge No.: 174

Date: 09/06/2015