

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01191 **Calib. Date:** 10/10/2014 **Calib. No.:** 00024
Certification File No.: 01192 **Cert. Date:** 10/10/2014 **Cert. No.:** 00016
Linearity File No.: 01193 **Lin. Date:** 10/10/2014 **Lin. No.:** 00016
Solution File No.: 01243 **Soln. Date:** 03/01/2015 **Soln. No.:** 00206
Sequential File No.: 01243 **File Date:** 03/01/2015

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDAB-0014
Control Solution %: 0.100% **Expires:** 04/29/2016
Solution Control Lot: 14D129 **Bottle No.:** 0150

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:17S		
Control 1 EC	0.100%	10:18S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:19S		
Control 2 EC	0.101%	10:19S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:20S		
Control 3 EC	0.101%	10:21S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:21S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:22S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRADY

First Name: PATRICK

MI: R

Signature:

PR #175

PR #175

Badge No.: 175

Date: 03/01/2015