

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01363 **Calib. Date:** 09/08/2015 **Calib. No.:** 00026
Certification File No.: 01364 **Cert. Date:** 09/08/2015 **Cert. No.:** 00018
Linearity File No.: 01365 **Lin. Date:** 09/08/2015 **Lin. No.:** 00018
Solution File No.: 01366 **Soln. Date:** 09/08/2015 **Soln. No.:** 00223
Sequential File No.: 01366 **File Date:** 09/08/2015

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDAB-0014
Control Solution %: 0.100% **Expires:** 08/22/2016
Solution Control Lot: 14H132 **Bottle No.:** 0055

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:23D		
Control 1 EC	0.100%	15:23D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.101%	15:23D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:24D		
Control 2 EC	0.100%	15:24D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	15:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:25D		
Control 3 EC	0.099%	15:26D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	15:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:26D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number:

DDXK P2-369 (C)

Changed By:

Last Name: DENNIS

First Name: MARC

MI: W.

Signature:

SGT. [Signature] #5925

Badge No.: 5925

Date: 09/08/2015