

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01363 **Calib. Date:** 09/08/2015 **Calib. No.:** 00026
Certification File No.: 01364 **Cert. Date:** 09/08/2015 **Cert. No.:** 00018
Linearity File No.: 01365 **Lin. Date:** 09/08/2015 **Lin. No.:** 00018
Solution File No.: 01371 **Soln. Date:** 09/25/2015 **Soln. No.:** 00224
Sequential File No.: 01371 **File Date:** 09/25/2015

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDAB-0014
Control Solution %: 0.100% **Expires:** 06/13/2017
Solution Control Lot: 15F139 **Bottle No.:** 0490

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:05D		
Control 1 EC	0.098%	11:06D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	11:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:06D		
Control 2 EC	0.097%	11:07D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.097%	11:07D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:07D		
Control 3 EC	0.096%	11:08D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	11:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:09D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL.

MI: A

Signature: _____

Badge No.: 129

Date: 09/25/2015