

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00995	Calib. Date:	09/08/2015
Certification File No.:	00996	Cert. Date:	09/08/2015
Linearity File No.:	00997	Lin. Date:	09/08/2015
Solution File No.:	01069	Soln. Date:	01/01/2016
Sequential File No.:	01069	File Date:	01/01/2016
		Calib. No.:	00019
		Cert. No.:	00016
		Lin. No.:	00016
		Soln. No.:	00202
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0015
Solution Control Lot:	15J149	Expires:	10/01/2017
		Bottle No.:	0951

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:28S		
Control 1 EC	0.100%	09:29S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.101%	09:29S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:30S		
Control 2 EC	0.100%	09:30S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:31S		
Control 3 EC	0.100%	09:32S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:32S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:32S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MALONE

First Name: MICHAEL

MI: T

Signature:  #153

Badge No.: 153

Date: 01/01/2016