

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01448 **Calib. Date:** 03/01/2016 **Calib. No.:** 00027
Certification File No.: 01449 **Cert. Date:** 03/01/2016 **Cert. No.:** 00019
Linearity File No.: 01450 **Lin. Date:** 03/01/2016 **Lin. No.:** 00019
Solution File No.: 01543 **Soln. Date:** 08/11/2016 **Soln. No.:** 00264
Sequential File No.: 01543 **File Date:** 08/11/2016

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDXB S3-113
Control Solution %: 0.100% **Expires:** 05/17/2018
Solution Control Lot: 201605E **Bottle No.:** 0033

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:47D		
Control 1 EC	0.101%	05:48D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	05:48D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:48D		
Control 2 EC	0.100%	05:49D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	05:49D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:49D		
Control 3 EC	0.100%	05:50D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	05:50D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:51D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SANTORO

First Name: DANIEL

MI: J

Signature: 

Badge No.: 165

Date: 08/11/2016