

# Calibrating Unit

## New Standard Solution Report

|                         |                       |              |             |
|-------------------------|-----------------------|--------------|-------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C | Serial No.:  | ARZL-0146   |
| Location:               | WALL TOWNSHIP POLICE  |              |             |
| Calibration File No.:   | 01254                 | Calib. Date: | 01/25/2017  |
| Certification File No.: | 01255                 | Cert. Date:  | 01/25/2017  |
| Linearity File No.:     | 01256                 | Lin. Date:   | 01/25/2017  |
| Solution File No.:      | 01275                 | Soln. Date:  | 03/09/2017  |
| Sequential File No.:    | 01275                 | File Date:   | 03/09/2017  |
| Calibrating Unit:       | WET                   | Model No.:   | CU-34       |
| Control Solution %:     | 0.100%                | Serial No.:  | DDXB S3-114 |
| Solution Control Lot:   | 16330                 | Expires:     | 10/28/2018  |
|                         |                       | Bottle No.:  | 0797        |

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 13:25S        |                               |                           |
| Control 1 EC      | 0.100%         | 13:26S        | 34.0°C                        | *** TEST PASSED ***       |
| Control 1 IR      | 0.099%         | 13:26S        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 13:26S        |                               |                           |
| Control 2 EC      | 0.100%         | 13:27S        | 34.0°C                        | *** TEST PASSED ***       |
| Control 2 IR      | 0.101%         | 13:27S        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 13:28S        |                               |                           |
| Control 3 EC      | 0.100%         | 13:28S        | 34.0°C                        | *** TEST PASSED ***       |
| Control 3 IR      | 0.099%         | 13:28S        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 13:29S        |                               |                           |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: \_\_\_\_\_

*PTL KS #129*

Badge No.: 129

Date: 03/09/2017