

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01254 **Calib. Date:** 01/25/2017 **Calib. No.:** 00026
Certification File No.: 01255 **Cert. Date:** 01/25/2017 **Cert. No.:** 00022
Linearity File No.: 01256 **Lin. Date:** 01/25/2017 **Lin. No.:** 00022
Solution File No.: 01280 **Soln. Date:** 03/23/2017 **Soln. No.:** 00237
Sequential File No.: 01280 **File Date:** 03/23/2017

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDXB S3-114
Control Solution %: 0.100% **Expires:** 10/28/2018
Solution Control Lot: 16330 **Bottle No.:** 0183

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:49S		
Control 1 EC	0.100%	07:49S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:50S		
Control 2 EC	0.100%	07:51S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	07:51S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:51S		
Control 3 EC	0.100%	07:52S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	07:52S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:53S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL.

MI: A

Signature: _____

PTL KAZANOWSKY #129

Badge No.: 129

Date: 03/23/2017

Completed 3/23/17