

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01254	Calib. Date:	01/25/2017
Certification File No.:	01255	Cert. Date:	01/25/2017
Linearity File No.:	01256	Lin. Date:	01/25/2017
Solution File No.:	01290	Soln. Date:	04/21/2017
Sequential File No.:	01290	File Date:	04/21/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-114
Solution Control Lot:	16330	Expires:	10/28/2018
		Bottle No.:	0182

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:13D		
Control 1 EC	0.101%	10:14D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:14D		
Control 2 EC	0.100%	10:15D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:16D		
Control 3 EC	0.100%	10:16D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:17D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____



Badge No.: 129

Date: 04/21/2017