

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	01613	Calib. Date: 01/25/2017
Certification File No.:	01614	Calib. No.: 00032
Linearity File No.:	01615	Cert. Date: 01/25/2017
Solution File No.:	01649	Cert. No.: 00023
Sequential File No.:	01649	Lin. Date: 01/25/2017
		Lin. No.: 00023
		Soln. Date: 05/18/2017
		Soln. No.: 00284
		File Date: 05/18/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-113
Solution Control Lot:	16330	Expires: 10/28/2018
		Bottle No.: 0062

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:29D		
Control 1 EC	0.100%	08:30D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	08:30D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:30D		
Control 2 EC	0.099%	08:31D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:31D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:32D		
Control 3 EC	0.098%	08:32D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:32D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:33D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PTL K #129

Badge No.: 129

Date: 05/18/2017