

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01613 Calib. Date: 01/25/2017 Calib. No.: 00032
Certification File No.: 01614 Cert. Date: 01/25/2017 Cert. No.: 00023
Linearity File No.: 01615 Lin. Date: 01/25/2017 Lin. No.: 00023
Solution File No.: 01657 Soln. Date: 06/02/2017 Soln. No.: 00285
Sequential File No.: 01657 File Date: 06/02/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-113
Control Solution %: 0.100% Expires: 10/28/2018
Solution Control Lot: 16330 Bottle No.: 0056

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:59D		
Control 1 EC	0.100%	09:59D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:59D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:00D		
Control 2 EC	0.098%	10:01D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:01D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:01D		
Control 3 EC	0.097%	10:02D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:02D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:03D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PK. K #129

Badge No.: 129

Date: 06/02/2017