

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0037  
**Location:** WALL TOWNSHIP POLICE  
**Calibration File No.:** 01613 **Calib. Date:** 01/25/2017 **Calib. No.:** 00032  
**Certification File No.:** 01614 **Cert. Date:** 01/25/2017 **Cert. No.:** 00023  
**Linearity File No.:** 01615 **Lin. Date:** 01/25/2017 **Lin. No.:** 00023  
**Solution File No.:** 01662 **Soln. Date:** 06/16/2017 **Soln. No.:** 00286  
**Sequential File No.:** 01662 **File Date:** 06/16/2017

**Calibrating Unit:** WET **Model No.:** CU-34 **Serial No.:** DDXB S3-113  
**Control Solution %:** 0.100% **Expires:** 10/28/2018  
**Solution Control Lot:** 16330 **Bottle No.:** 0850

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 16:30D        |                               |                           |
| Control 1 EC      | 0.099%         | 16:30D        | 33.9°C                        | *** TEST PASSED ***       |
| Control 1 IR      | 0.097%         | 16:30D        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 16:31D        |                               |                           |
| Control 2 EC      | 0.097%         | 16:32D        | 33.9°C                        | *** TEST PASSED ***       |
| Control 2 IR      | 0.097%         | 16:32D        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 16:32D        |                               |                           |
| Control 3 EC      | 0.096%         | 16:33D        | 33.9°C                        | *** TEST PASSED ***       |
| Control 3 IR      | 0.099%         | 16:33D        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 16:34D        |                               |                           |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: \_\_\_\_\_

*PK #129*

Badge No.: 129

Date: 06/16/2017